



**Abstract:**

BD/CareFusion 303 is recalling the Alaris™ PC unit and PC Unit Front Case with Keypad Replacement Kits because the keypad may have one or more keys in the keypad that become unresponsive or stuck. This could lead to an infusion delay or prevent clinicians from changing fluid or medication infusions on the affected device.

Use of the affected product may cause serious adverse events, including death.

There have been 1,186 complaints about this device issue and one reported injury. There have been no reported deaths.

[Full Alert](#)