



...voluntary recall on 29 July 2019 of certain lots of Plum and Sapphire Microbore Infusion Sets with inline filters due to the potential for small amounts of fluid leaking out of the air vents on the inline filters” FDA (2019).

FDA report “ICU Medical, Inc. announced a voluntary recall on 29 July 2019 of certain lots of Plum and Sapphire Microbore Infusion Sets with inline filters due to the potential for small amounts of fluid leaking out of the air vents on the inline filters. To date, ICU Medical is not aware of adverse events related to this matter and is issuing this notification out of an abundance of caution.

You may also be interested in...

FDA report of single lot of IV fluid voluntary recall by ICU Medical
State of needlestick training for undergraduate medical students
Medical device classification in vascular access

Full Alert

