



Becton Dickinson has issued a product recall against Vacutainer plus heparin tube ref 367885 lot number 5084266 expiry date 2016-08" NHS Supply Chain (2015).

NHS Supply Chain report "Becton Dickinson has issued a product recall against Vacutainer plus heparin tube ref 367885 lot number 5084266 expiry date 2016-08. This follows discovery that a small number of these tubes contain sodium fluoride / potassium oxalate, an inhibitor of the glycolytic pathway, instead of the specified lithium heparin additive."

ReTweet if useful... Becton Dickinson product recall against Vacutainer plus heparin tube <http://ctt.ec/8Kb3d+> @ivteam #ivteam

Click To Tweet

[Read Full Alert](#)

**Thank you to our partners for supporting IVTEAM**

