



OriGen Biomedical initiated a nationwide recall for two (2) lots of VV28F Reinforced Dual Lumen ECMO Catheters” FDA (2017).

FDA report “OriGen Biomedical initiated a nationwide recall for two (2) lots of VV28F Reinforced Dual Lumen ECMO Catheters. These VV28F Reinforced Dual Lumen ECMO Catheters have been found to have the potential for a separation of the clear extension tube from the hub that it is inserted in, which potentially could result in required intervention to prevent permanent impairment/damage.”

Full Alert

ReTweet if useful... OriGen Biomedical initiated a nationwide recall for reinforced dual lumen ECMO catheters <https://ctt.ec/VWRfe+> @ivteam #ivteam

Click To Tweet

Thank you to our partners for supporting IVTEAM

