



Intravenous products: The MHRA have issued a Medical Device Alert that relates to all Plum A+ infusion pumps manufactured by Hospira Inc. The alert 'MHRA Safety Alert MDA/2011/110' highlights a "Possible delay in therapy due to continuous recycling and/or rebooting if the LED display settings are adjusted from the original default settings."

[Click here to view the alert.](#)

