



Nurse Assist announced voluntary recall of all unexpired lots of I.V. Flush Syringes due to a potential link to Burkholderia cepacia bloodstream infections with the product” FDA (2016).

FDA report “Nurse Assist announced voluntary recall of all unexpired lots of I.V. Flush Syringes due to a potential link to Burkholderia cepacia bloodstream infections with the product. According to the U.S. Centers for Disease Control and Prevention (CDC), the effects of Burkholderia cepacia on people “vary widely, ranging from no symptoms at all to serious respiratory infections, especially in patients with cystic fibrosis.” Nurse Assist voluntarily recalled its I.V. Flush Syringes after becoming aware of patients that developed Burkholderia cepacia bloodstream infections while receiving intravenous care using prepackaged saline flushes from Nurse Assist.”

ReTweet if useful... Recall of IV flush syringes by Nurse Assist [@ivteam](http://ctt.ec/ERLda+) #ivteam

Click To Tweet

Full Alert

Thank you to our partners for supporting IVTEAM

